

## Hatch Sure Calibration Certificate

This Hatch Leak Testing Device has been calibrated using the following equipment.  
This equipment has been calibrated and is traceable to national standards.

TG101A Serial No. 312959

TDS3014 Serial No. B010408

Calibration No.	371	Receiver SN:	266
Calibration Date:	21/11/2017	Transmitter SN:	266
Calibrated By	Matt Winter	Microphone 1 W.O No:	W-10957
Order No:	S023555	Microphone 2 W.O No:	W-10957

Calibrated For	Navitech Shipping Service
----------------	---------------------------

### Calibration Results

Source dB	Measured dB	Acceptance Limit	Result (Pass/Fail)
0dB	0	+/- 1dB	Pass
20dB	20	+/- 1dB	Pass
40dB	40	+/- 1dB	Pass
60dB	60	+/- 1dB	Pass

Transmitter Elements Tested?

OH Calibration Tested?

All other items inspected?

Inspected By: Matt Winter

Calibration Due Date: 21 November 2019

Calibrated By Signature & Stamp:

Inspected By Signature & Stamp:



*"Simplicity through technology"*

